
Midmark M11 - v1.0.0
Total Cycles: 5490

Name: RITTER 1
Sterilizer ID: V2178198
04/09/2025 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 94.3 F 0.3
4:00 88.9 F 1.9
6:00 129.5 F 4.4
8:00 152.3 F 7.2
10:00 176.9 F 10.6
12:00 200.5 F 4.6
14:00 210.1 F 4.6
16:00 221.2 F 7.9
18:00 231.0 F 11.7
20:00 229.8 F 6.7
22:00 241.8 F 11.7
24:00 237.2 F 8.9
26:00 249.9 F 15.0
28:00 260.5 F 21.0
30:00 270.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.3
2:00 272.4 F 28.7
3:00 272.2 F 28.7
4:00 271.9 F 28.1

Min 271.5 F 28.1
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 197.5 F 0.0
10:00 193.1 F 0.0
15:00 195.0 F 0.0
20:00 203.5 F 0.0
25:00 212.6 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 30:44
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:09:02
CYCLE COMPLETE

04/09/2025 10:06 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____