

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005535  
14:21 15/05/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:24  
PRESSURIZING 2:26  
STERILIZING 3:20  
136.4 C 223kPa 3:20  
136.4 C 222kPa 3:50  
136.1 C 222kPa 4:20  
136.4 C 225kPa 4:50  
136.4 C 219kPa 5:20  
136.2 C 222kPa 5:50  
136.2 C 221kPa 6:20  
136.4 C 223kPa 6:50  
136.4 C 223kPa 6:50  
VENTING 6:50  
AIR DRYING 7:43  
CYCLE COMPLETE 10:46

-----  
Digital Signature #  
C3681E4FEA8EB9A2  
-----