
Midmark M11 - v1.0.0
Total Cycles: 1095

Name: RITTER 1
Sterilizer ID: V2178198
08/19/2021 06:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.4 F 0.0
2:00 77.7 F 0.0
4:00 102.4 F 0.0
6:00 136.6 F 0.1
8:00 157.4 F 0.1
10:00 184.2 F 0.3
12:00 203.0 F 3.3
14:00 213.9 F 6.3
16:00 223.2 F 9.6
18:00 233.2 F 13.3
20:00 233.6 F 8.7
22:00 243.3 F 12.4
24:00 242.2 F 11.5
26:00 239.3 F 9.9
28:00 252.9 F 16.5
30:00 264.7 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.9
1:00 272.4 F 28.3
2:00 272.2 F 28.6
3:00 271.8 F 28.0
4:00 272.7 F 28.8

Min 271.5 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.3 F 0.0
5:00 194.0 F 0.0
10:00 194.3 F 0.0
15:00 191.9 F 0.0
20:00 186.6 F 0.0
25:00 183.4 F 0.0
30:00 178.6 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 31:51
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:10:09
CYCLE COMPLETE

08/19/2021 08:00 AM
ACCEPTED: X
REJECTED:
Operator: U02