
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000546
15:22 09/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:42
STERILIZING 3:48
136.0 C 220kPa 3:48
136.4 C 233kPa 4:18
136.2 C 231kPa 4:48
136.3 C 232kPa 5:18
136.3 C 232kPa 5:48
136.5 C 233kPa 6:18
136.3 C 232kPa 6:48
136.4 C 233kPa 7:18
136.4 C 233kPa 7:18
VENTING 7:18
AIR DRYING 8:12
CYCLE COMPLETE 19:58

Digital Signature #
FA5C8BEE05B23636
