
Midmark M11 - v1.0.0
Total Cycles: 2827

Name: RITTER 1
Sterilizer ID: V2178198
01/30/2023 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.7 F 0.0
2:00 196.8 F 4.7
4:00 198.0 F 9.6
6:00 217.8 F 6.1
8:00 228.5 F 9.4
10:00 235.3 F 9.0
12:00 242.6 F 12.5
14:00 241.9 F 11.3
16:00 255.2 F 18.2
18:00 267.0 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.4 F 28.8
2:00 272.3 F 29.1
3:00 271.9 F 28.5
4:00 272.6 F 29.2

Min 271.6 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 213.3 F 0.0
10:00 228.4 F 0.0
15:00 227.0 F 0.0
20:00 223.8 F 0.0
25:00 219.1 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:34
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:13
CYCLE COMPLETE

01/30/2023 03:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____