
Midmark M11 - v1.0.0
Total Cycles: 975

Name: RITTER 1
Sterilizer ID: V2178198
07/07/2021 11:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.8 F 0.0
2:00 187.5 F 2.8
4:00 194.5 F 6.1
6:00 218.0 F 10.3
8:00 225.1 F 5.4
10:00 231.7 F 8.9
12:00 242.2 F 12.2
14:00 239.5 F 10.9
16:00 237.3 F 9.0
18:00 250.2 F 15.2
20:00 261.1 F 21.4
22:00 270.9 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 29.1
2:00 272.2 F 28.6
3:00 271.4 F 28.4
4:00 272.5 F 29.2

Min 271.4 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.2 F 0.0
5:00 229.4 F 0.0
10:00 230.5 F 0.0
15:00 230.4 F 0.0
20:00 229.5 F 0.0
25:00 229.0 F 0.0
30:00 230.6 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 22:47
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:01:28
CYCLE COMPLETE

07/07/2021 12:14 PM
ACCEPTED: X
REJECTED:
Operator: U01