
Midmark M11 - v1.0.5
Total Cycles: 2337

Name: RITTER2
Sterilizer ID: V2313228
05/04/2023 08:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.2 F 0.0
2:00 158.1 F 2.6
4:00 167.5 F 5.7
6:00 193.6 F 10.4
8:00 212.1 F 4.8
10:00 227.1 F 10.0
12:00 231.1 F 8.1
14:00 241.8 F 12.9
16:00 241.3 F 11.7
18:00 254.2 F 18.3
20:00 265.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.4 F 29.6
2:00 271.7 F 29.7
3:00 272.3 F 29.5
4:00 271.6 F 29.7

Min 271.3 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 211.7 F 0.0
10:00 209.9 F 0.0
15:00 224.3 F 0.0
20:00 235.2 F 0.0
25:00 228.3 F 0.0
30:00 218.8 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 21:41
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

05/04/2023 09:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____