
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 001852
13:45 15/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:22
PRESSURIZING 2:27
STERILIZING 3:30
136.0 C 222kPa 3:30
136.2 C 224kPa 4:00
136.3 C 225kPa 4:30
136.4 C 226kPa 5:00
136.3 C 224kPa 5:30
136.3 C 225kPa 6:00
136.4 C 226kPa 6:30
136.3 C 226kPa 7:00
136.3 C 226kPa 7:00
VENTING 7:00
AIR DRYING 7:54
CYCLE COMPLETE 8:09

Digital Signature #
D5AE6D2746064D66
