
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000544
13:11 09/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:46
STERILIZING 3:52
136.0 C 228kPa 3:52
136.4 C 233kPa 4:22
136.2 C 231kPa 4:52
136.1 C 229kPa 5:22
136.4 C 233kPa 5:52
136.3 C 232kPa 6:22
136.2 C 232kPa 6:52
136.1 C 231kPa 7:22
136.1 C 231kPa 7:22
VENTING 7:22
AIR DRYING 8:12
CYCLE COMPLETE 23:12

Digital Signature #
3E4DC8236C85D33E
