
Midmark M11 - v1.0.0
Total Cycles: 954

Name: RITTER 1
Sterilizer ID: V2178198
06/29/2021 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.3 F 0.0
2:00 136.7 F 1.1
4:00 155.8 F 4.1
6:00 179.9 F 8.6
8:00 209.9 F 8.8
10:00 234.7 F 8.8
12:00 233.0 F 9.4
14:00 240.9 F 11.5
16:00 238.0 F 9.2
18:00 251.5 F 15.7
20:00 262.9 F 22.4
22:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 28.6
1:00 272.2 F 28.8
2:00 271.5 F 28.9
3:00 271.9 F 28.5
4:00 272.1 F 28.8

Min 271.3 F 28.5
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 174.3 F 0.0
5:00 238.0 F 0.0
10:00 234.0 F 0.0
15:00 226.1 F 0.0
20:00 221.3 F 0.0
25:00 209.4 F 0.0
30:00 220.8 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 22:34
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 01:00:41
CYCLE COMPLETE

06/29/2021 01:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____