

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3312

Name: RITTER 1  
Sterilizer ID: V2178198  
05/31/2023 10:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 205.0 F 0.0  
2:00 198.4 F 1.2  
4:00 223.8 F 6.9  
6:00 233.3 F 12.4  
8:00 233.6 F 9.0  
10:00 240.8 F 9.9  
12:00 245.8 F 13.4  
14:00 244.9 F 12.5  
16:00 258.0 F 19.5  
18:00 269.4 F 26.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.1  
1:00 272.5 F 28.6  
2:00 271.7 F 28.3  
3:00 272.5 F 28.6  
4:00 271.6 F 28.0  
  
Min 271.6 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.2 F 0.0  
5:00 225.2 F 0.0  
10:00 231.4 F 0.0  
15:00 235.2 F 0.0  
20:00 228.0 F 0.0  
25:00 223.1 F 0.0  
30:00 219.8 F 0.0  
Drying Complete

FILLING: 2:00  
HEATING: 19:00  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:13  
CYCLE COMPLETE

05/31/2023 11:36 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_