

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4960

Name: RITTER 1  
Sterilizer ID: V2178198  
10/03/2024 12:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.4 F 0.0  
2:00 161.0 F 1.8  
4:00 186.7 F 5.7  
6:00 204.8 F 10.4  
8:00 223.9 F 4.8  
10:00 228.3 F 9.5  
12:00 237.0 F 9.2  
14:00 242.4 F 12.4  
16:00 240.5 F 10.8  
18:00 253.6 F 17.3  
20:00 264.9 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.4 F 29.0  
2:00 271.3 F 28.4  
3:00 272.4 F 29.1  
4:00 271.6 F 28.5

Min 271.2 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.4 F 0.0  
5:00 213.4 F 0.0  
10:00 226.4 F 0.0  
15:00 233.3 F 0.0  
20:00 227.9 F 0.0  
25:00 225.0 F 0.0  
30:00 235.2 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 21:53  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:00:57  
CYCLE COMPLETE

10/03/2024 01:45 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_