
Midmark M11 - v1.0.0
Total Cycles: 2811

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2023 10:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 202.4 F 0.0
2:00 199.1 F 1.5
4:00 198.2 F 6.3
6:00 216.8 F 11.7
8:00 226.0 F 7.4
10:00 239.5 F 13.0
12:00 239.8 F 11.0
14:00 238.3 F 9.7
16:00 253.1 F 16.9
18:00 265.3 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.1
2:00 271.8 F 28.3
3:00 272.3 F 29.0
4:00 272.6 F 28.9

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 230.1 F 0.0
10:00 243.8 F 0.0
15:00 232.1 F 0.0
20:00 227.1 F 0.0
25:00 228.4 F 0.0
30:00 229.4 F 0.0

Drying Complete

FILLING: 1:31
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

01/24/2023 11:51 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____