

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1593

Name: RITTER2  
Sterilizer ID: V2313228  
10/13/2022 10:18 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.6 F 0.0  
2:00 169.3 F 2.4  
4:00 185.3 F 6.6  
6:00 197.6 F 11.8  
8:00 225.7 F 7.1  
10:00 234.0 F 12.5  
12:00 236.1 F 9.7  
14:00 235.8 F 9.0  
16:00 247.1 F 14.5  
18:00 259.4 F 21.2  
20:00 270.1 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.9  
1:00 273.1 F 30.4  
2:00 271.5 F 30.1  
3:00 273.0 F 30.4  
4:00 271.4 F 29.9

Min 271.3 F 29.8  
Max 273.1 F 30.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.1 F 0.0  
5:00 236.8 F 0.0  
10:00 229.0 F 0.0  
15:00 228.5 F 0.0  
20:00 229.2 F 0.0  
25:00 229.1 F 0.0  
30:00 220.1 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 21:49  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:00:46  
CYCLE COMPLETE

10/13/2022 11:19 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02