

-----  
Midmark M11 - v1.0.5  
Total Cycles: 973

Name: RITTER2  
Sterilizer ID: V2313228  
04/11/2022 09:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.6 F 0.0  
2:00 192.5 F 0.7  
4:00 196.0 F 5.4  
6:00 215.4 F 10.7  
8:00 225.6 F 6.6  
10:00 238.9 F 12.5  
12:00 239.3 F 10.8  
14:00 238.7 F 10.1  
16:00 253.4 F 17.4  
18:00 265.7 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.8  
1:00 272.5 F 29.5  
2:00 271.7 F 28.7  
3:00 272.2 F 29.6  
4:00 272.2 F 29.2  
  
Min 271.3 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.4 F 0.0  
5:00 232.7 F 0.0  
10:00 222.0 F 0.0  
15:00 222.3 F 0.0  
20:00 220.4 F 0.0  
25:00 220.6 F 0.0  
30:00 219.3 F 0.0  
Drying Complete

FILLING: 1:34  
HEATING: 19:40  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:58:23  
CYCLE COMPLETE

04/11/2022 10:55 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_