

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001526  
10:24 10/22/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 5:01  
PRESSURIZING 6:49  
STERILIZING 8:09  
136.0 C 220kPa 8:09  
136.5 C 216kPa 8:39  
135.9 C 218kPa 9:09  
136.2 C 221kPa 9:39  
136.3 C 221kPa 10:09  
135.8 C 219kPa 10:39  
136.4 C 222kPa 11:09  
136.4 C 219kPa 11:39  
VENTING 11:39  
AIR DRYING 12:41  
CYCLE COMPLETE 32:00

-----  
Digital Signature #  
0B164B22308A330F  
-----