

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2467

Name: RITTER 1  
Sterilizer ID: V2178198  
10/21/2022 07:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 77.7 F 0.0  
2:00 86.9 F 0.0  
4:00 100.4 F 0.0  
6:00 114.3 F 2.0  
8:00 146.8 F 4.6  
10:00 170.5 F 7.7  
12:00 192.3 F 11.7  
14:00 210.4 F 3.5  
16:00 222.9 F 7.3  
18:00 234.8 F 11.2  
20:00 232.6 F 6.6  
22:00 242.2 F 12.5  
24:00 239.9 F 10.3  
26:00 252.1 F 16.5  
28:00 263.3 F 23.0  
30:00 272.2 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.4  
1:00 272.6 F 29.2  
2:00 271.6 F 28.4  
3:00 272.0 F 29.0  
4:00 272.4 F 28.9

Min 271.1 F 28.2  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.2 F 0.0  
5:00 202.3 F 0.0  
10:00 192.3 F 0.0  
15:00 194.7 F 0.0  
20:00 195.8 F 0.0  
25:00 213.5 F 0.0  
30:00 227.8 F 0.0  
Drying Complete

FILLING: 0:44  
HEATING: 30:14  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:08:09  
CYCLE COMPLETE

10/21/2022 08:35 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_