
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001832
7:47 22/06/2021

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:41
STERILIZING 4:52
136.4 C 223kPa 4:52
136.3 C 223kPa 5:23
136.4 C 222kPa 5:53
136.4 C 223kPa 6:23
136.3 C 223kPa 6:53
136.3 C 222kPa 7:23
136.4 C 224kPa 7:53
136.2 C 222kPa 8:23
136.2 C 222kPa 8:23
VENTING 8:23
AIR DRYING 9:16
CYCLE COMPLETE 9:31

Digital Signature #
1406992E8CE31E70
