
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076

Unit # : 000

WATER QUALITY

2.3uS / 1.4ppm

CYCLE NUMBER 000010

11:15 13/01/2020

RUBBER AND PLASTICS

121 C for 15 min.

CYCLE START 0:00

CONDITIONING 4:04

PRESSURIZING 9:47

STERILIZING 10:43

123.0 C 117kPa 10:43

123.5 C 120kPa 11:13

123.4 C 120kPa 11:43

123.4 C 120kPa 12:13

123.4 C 120kPa 12:43

123.4 C 120kPa 13:13

123.5 C 120kPa 13:43

123.5 C 120kPa 14:14

123.5 C 120kPa 14:44

123.5 C 120kPa 15:14

123.5 C 120kPa 15:44

123.5 C 120kPa 16:14

123.5 C 120kPa 16:44

123.4 C 120kPa 17:14

123.5 C 120kPa 17:44

123.5 C 120kPa 18:14

123.4 C 120kPa 18:44

123.1 C 120kPa 19:14

123.4 C 120kPa 19:44

123.4 C 120kPa 20:14

123.5 C 120kPa 20:44

123.5 C 120kPa 21:14

123.4 C 120kPa 21:44

123.5 C 121kPa 22:14

123.4 C 120kPa 22:44

123.4 C 120kPa 23:14

123.4 C 120kPa 23:44

123.3 C 120kPa 24:14

123.5 C 120kPa 24:44

123.4 C 120kPa 25:14

123.5 C 121kPa 25:43

VENTING 25:43

AIR DRYING 26:46

CYCLE COMPLETE 56:47

Digital Signature #

EA8EB45EE7C60CE7
