
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000580
12:07 12/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
CYCLE ABORTED 2:47
WATER LEVEL LOW
NOT STERILE
END TIME 2:47

Digital Signature #
CE757577FC396D0C
