
Midmark M11 - v1.0.5
Total Cycles: 3195

Name: RITTER2
Sterilizer ID: V2313228
12/07/2023 01:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.9 F 0.0
2:00 104.0 F 0.1
4:00 111.9 F 1.0
6:00 139.4 F 3.5
8:00 164.1 F 7.0
10:00 187.0 F 11.1
12:00 216.1 F 4.4
14:00 221.1 F 8.5
16:00 231.6 F 13.1
18:00 234.0 F 9.8
20:00 236.7 F 9.4
22:00 245.7 F 13.9
24:00 257.3 F 20.1
26:00 268.0 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.5 F 30.1
2:00 272.6 F 29.7
3:00 271.4 F 29.1
4:00 271.4 F 29.6

Min 271.1 F 29.1
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.5 F 0.0
5:00 213.2 F 0.0
10:00 209.4 F 0.0
15:00 196.6 F 0.0
20:00 191.2 F 0.0
25:00 186.9 F 0.0
30:00 185.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 27:25
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:05:38
CYCLE COMPLETE

12/07/2023 02:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____