
Midmark M11 - v1.0.5
Total Cycles: 4399

Name: RITTER2
Sterilizer ID: V2313228
03/05/2025 09:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.4 F 0.0
2:00 71.8 F 0.2
4:00 73.5 F 0.9
6:00 103.2 F 2.8
8:00 144.7 F 6.2
10:00 169.5 F 10.0
12:00 184.5 F 9.8
14:00 206.4 F 3.9
16:00 217.2 F 7.1
18:00 227.3 F 10.8
20:00 229.7 F 7.7
22:00 239.1 F 11.7
24:00 237.0 F 9.7
26:00 249.6 F 15.8
28:00 260.5 F 21.9
30:00 270.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.4 F 29.5
2:00 272.8 F 30.0
3:00 271.8 F 29.3
4:00 272.6 F 29.6

Min 271.7 F 29.1
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 201.2 F 0.0
10:00 195.8 F 0.0
15:00 193.7 F 0.0
20:00 190.9 F 0.0
25:00 190.3 F 0.0
30:00 199.1 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:48
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:09:10
CYCLE COMPLETE

03/05/2025 10:47 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____