
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002362
9:56 06/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:02
STERILIZING 3:02
136.4 C 224kPa 3:02
136.3 C 226kPa 3:33
136.2 C 223kPa 4:03
136.4 C 226kPa 4:33
136.4 C 224kPa 5:03
136.1 C 223kPa 5:33
136.3 C 224kPa 6:03
136.4 C 224kPa 6:33
136.4 C 224kPa 6:33
VENTING 6:33
AIR DRYING 7:26
CYCLE COMPLETE 37:26

Digital Signature #
33AB6E1951A01CC3
