

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3352

Name: RITTER 1  
Sterilizer ID: V2178198  
06/07/2023 01:51 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 157.0 F 0.0  
2:00 161.9 F 0.0  
4:00 188.9 F 4.1  
6:00 210.6 F 9.1  
8:00 223.3 F 8.2  
10:00 231.7 F 9.7  
12:00 234.8 F 8.9  
14:00 243.6 F 12.9  
16:00 242.4 F 11.7  
18:00 255.4 F 18.3  
20:00 266.7 F 25.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.4  
1:00 272.8 F 29.3  
2:00 272.3 F 28.8  
3:00 271.3 F 28.3  
4:00 272.3 F 29.1

Min 271.3 F 28.3  
Max 273.1 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.8 F 0.0  
5:00 217.1 F 0.0  
10:00 239.4 F 0.0  
15:00 226.0 F 0.0  
20:00 222.9 F 0.0  
25:00 221.5 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 21:34  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:59:50  
CYCLE COMPLETE

06/07/2023 02:50 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_