
Midmark M11 - v1.0.5
Total Cycles: 298

Name: RITTER2
Sterilizer ID: V2313228
08/23/2021 06:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.5 F 0.0
2:00 82.7 F 0.0
4:00 124.7 F 0.1
6:00 106.0 F 0.2
8:00 144.1 F 2.5
10:00 169.9 F 5.6
12:00 205.2 F 9.6
14:00 207.8 F 9.2
16:00 226.3 F 7.4
18:00 236.6 F 11.7
20:00 232.8 F 8.3
22:00 243.3 F 12.3
24:00 244.1 F 12.7
26:00 256.1 F 19.1
28:00 266.5 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.7 F 29.4
2:00 272.2 F 29.6
3:00 272.1 F 29.1
4:00 272.6 F 29.8

Min 271.6 F 28.9
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 236.8 F 0.0
10:00 220.5 F 0.0
15:00 218.9 F 0.0
20:00 213.3 F 0.0
25:00 216.1 F 0.0
30:00 209.9 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 29:33
STERILIZING: 4:00
VENTING: 2:60
DRYING: 30:00
TOTAL CYCLE: 01:07:32
CYCLE COMPLETE

08/23/2021 08:02 AM
ACCEPTED: X
REJECTED:
Operator: U02