
Midmark M11 - v1.0.5
Total Cycles: 927

Name: RITTER2
Sterilizer ID: V2313228
03/30/2022 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.5 F 0.0
2:00 188.1 F 0.6
4:00 198.2 F 6.0
6:00 220.6 F 11.7
8:00 230.2 F 8.7
10:00 238.0 F 10.1
12:00 244.0 F 12.4
14:00 246.9 F 13.7
16:00 260.2 F 21.1
18:00 272.0 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.6
1:00 272.6 F 29.5
2:00 271.9 F 28.8
3:00 271.1 F 28.8
4:00 272.5 F 29.5

Min 271.0 F 28.5
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 232.1 F 0.0
10:00 227.3 F 0.0
15:00 222.0 F 0.0
20:00 222.3 F 0.0
25:00 224.0 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 18:35
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:06
CYCLE COMPLETE

03/30/2022 10:46 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____