
Midmark M11 - v1.0.5
Total Cycles: 4188

Name: RITTER2
Sterilizer ID: V2313228
12/05/2024 11:21 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.5 F 0.0
2:00 166.1 F 0.4
4:00 179.4 F 2.9
6:00 208.3 F 7.4
8:00 215.0 F 12.6
10:00 227.3 F 8.0
12:00 238.5 F 13.1
14:00 238.7 F 11.1
16:00 237.5 F 10.0
18:00 251.3 F 16.7
20:00 262.7 F 23.3
22:00 272.8 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.7 F 30.0
2:00 271.7 F 29.3
3:00 271.6 F 29.7
4:00 272.4 F 29.8

Min 271.1 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 211.0 F 0.0
10:00 213.7 F 0.0
15:00 219.9 F 0.0
20:00 206.9 F 0.0
25:00 213.3 F 0.0
30:00 210.5 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:01:22
CYCLE COMPLETE

12/05/2024 12:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____