
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 003240
16:56 28/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:46
PRESSURIZING 2:53
STERILIZING 3:59
136.4 C 220kPa 3:59
136.4 C 224kPa 4:29
136.3 C 225kPa 4:59
136.4 C 225kPa 5:29
136.4 C 225kPa 5:59
136.1 C 223kPa 6:29
136.4 C 226kPa 6:59
136.3 C 224kPa 7:29
136.3 C 224kPa 7:29
VENTING 7:29
AIR DRYING 8:22
CYCLE COMPLETE 8:28

Digital Signature #
E26731DABB821A6C
