

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3989

Name: RITTER2  
Sterilizer ID: V2313228  
10/02/2024 01:30 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.4 F 0.0  
2:00 124.0 F 0.4  
4:00 135.7 F 2.8  
6:00 167.1 F 6.2  
8:00 189.9 F 10.3  
10:00 218.9 F 7.2  
12:00 223.3 F 8.5  
14:00 235.4 F 12.8  
16:00 235.0 F 9.3  
18:00 240.6 F 10.8  
20:00 243.5 F 12.4  
22:00 254.4 F 18.0  
24:00 264.0 F 23.8  
26:00 272.5 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.5  
1:00 271.8 F 29.8  
2:00 271.6 F 30.1  
3:00 272.2 F 30.2  
4:00 272.6 F 30.1

Min 270.9 F 29.5  
Max 273.3 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.2 F 0.0  
5:00 225.8 F 0.0  
10:00 214.4 F 0.0  
15:00 222.0 F 0.0  
20:00 176.0 F 0.0  
25:00 168.9 F 0.0  
30:00 168.1 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 26:56  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:05:38  
CYCLE COMPLETE

10/02/2024 02:35 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_