

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2855

Name: RITTER2  
Sterilizer ID: V2313228  
09/08/2023 02:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.3 F 0.0  
2:00 178.3 F 3.4  
4:00 194.7 F 8.1  
6:00 212.3 F 13.1  
8:00 221.6 F 8.1  
10:00 233.9 F 13.1  
12:00 237.6 F 10.7  
14:00 236.2 F 9.3  
16:00 249.7 F 15.7  
18:00 260.8 F 22.1  
20:00 270.9 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.1  
1:00 272.4 F 29.9  
2:00 271.7 F 29.0  
3:00 272.6 F 29.8  
4:00 271.8 F 29.2

Min 271.3 F 29.0  
Max 272.6 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.9 F 0.0  
5:00 204.6 F 0.0  
10:00 201.6 F 0.0  
15:00 197.7 F 0.0  
20:00 193.3 F 0.0  
25:00 189.5 F 0.0  
30:00 192.8 F 0.0  
Drying Complete

FILLING: 2:10  
HEATING: 20:36  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:00:19  
CYCLE COMPLETE

09/08/2023 03:56 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_