

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2760

Name: RITTER 1  
Sterilizer ID: V2178198  
01/10/2023 03:12 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
2:00 190.8 F 0.0  
4:00 205.9 F 5.3  
6:00 219.9 F 10.7  
8:00 226.3 F 6.4  
10:00 239.7 F 12.3  
12:00 239.2 F 10.4  
14:00 237.2 F 9.1  
16:00 252.6 F 16.4  
18:00 264.8 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.6 F 28.9  
2:00 271.8 F 28.1  
3:00 272.4 F 29.0  
4:00 272.3 F 28.6  
  
Min 271.4 F 28.0  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.2 F 0.0  
5:00 223.4 F 0.0  
10:00 218.6 F 0.0  
15:00 239.7 F 0.0  
20:00 241.9 F 0.0  
25:00 230.7 F 0.0  
30:00 222.1 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 19:51  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:58  
CYCLE COMPLETE

01/10/2023 04:09 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_