
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
2.1uS / 1.3ppm
CYCLE NUMBER 003734
10:35 31/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:41
STERILIZING 4:31
136.4 C 222kPa 4:31
136.2 C 223kPa 5:01
136.4 C 224kPa 5:31
136.4 C 223kPa 6:01
136.3 C 221kPa 6:31
136.3 C 221kPa 7:01
136.4 C 221kPa 7:31
136.3 C 222kPa 8:01
136.3 C 222kPa 8:01
VENTING 8:02
AIR DRYING 9:18
CYCLE COMPLETE 9:22

Digital Signature #
A9179333D71F3D9B
