
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002398
9:56 08/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:56
STERILIZING 2:57
136.4 C 223kPa 2:57
136.2 C 222kPa 3:27
136.2 C 224kPa 3:57
136.4 C 224kPa 4:27
136.3 C 223kPa 4:57
136.4 C 225kPa 5:27
136.2 C 222kPa 5:57
136.4 C 219kPa 6:27
136.4 C 219kPa 6:27
VENTING 6:27
AIR DRYING 7:20
CYCLE COMPLETE 7:25

Digital Signature #
BB92989E3908EE43
