
Midmark M11 - v1.0.0
Total Cycles: 3970

Name: RITTER 1
Sterilizer ID: V2178198
11/02/2023 08:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 77.0 F 0.3
4:00 88.1 F 1.2
6:00 126.6 F 3.5
8:00 147.7 F 6.3
10:00 170.9 F 10.0
12:00 207.3 F 6.3
14:00 213.9 F 4.9
16:00 224.8 F 8.3
18:00 234.8 F 12.1
20:00 231.5 F 7.3
22:00 242.7 F 12.4
24:00 239.8 F 10.0
26:00 252.3 F 16.2
28:00 263.2 F 22.6
30:00 272.4 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.7 F 28.7
2:00 271.4 F 27.9
3:00 272.6 F 28.9
4:00 272.1 F 28.2

Min 271.3 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 223.2 F 0.0
10:00 232.0 F 0.0
15:00 230.9 F 0.0
20:00 222.2 F 0.0
25:00 222.0 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:14
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:08:27
CYCLE COMPLETE

11/02/2023 09:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____