

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3594

Name: RITTER 1  
Sterilizer ID: V2178198  
07/31/2023 10:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.4 F 0.0  
2:00 174.1 F 0.0  
4:00 175.9 F 1.0  
6:00 198.7 F 5.0  
8:00 210.4 F 9.3  
10:00 224.1 F 12.6  
12:00 230.3 F 8.3  
14:00 241.1 F 13.2  
16:00 240.2 F 10.8  
18:00 237.1 F 9.1  
20:00 251.1 F 15.6  
22:00 262.4 F 22.1  
24:00 272.7 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.0  
1:00 272.4 F 28.8  
2:00 271.8 F 28.1  
3:00 272.1 F 28.7  
4:00 272.2 F 28.5

Min 271.4 F 27.9  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.8 F 0.0  
5:00 214.3 F 0.0  
10:00 197.0 F 0.0  
15:00 194.9 F 0.0  
20:00 191.3 F 0.0  
25:00 188.0 F 0.0  
30:00 187.1 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 24:16  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:02:40  
CYCLE COMPLETE

07/31/2023 12:01 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_