
Midmark M11 - v1.0.0
Total Cycles: 2245

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2022 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.0 F 0.0
2:00 199.1 F 0.2
4:00 197.8 F 1.0
6:00 217.1 F 4.5
8:00 231.1 F 9.1
10:00 239.1 F 10.9
12:00 239.2 F 10.9
14:00 236.4 F 8.6
16:00 240.3 F 9.9
18:00 248.5 F 14.2
20:00 260.5 F 20.9
22:00 271.2 F 27.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.6 F 28.6
2:00 271.7 F 28.2
3:00 272.5 F 28.5
4:00 271.5 F 28.0

Min 271.5 F 27.9
Max 273.0 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.0 F 0.0
5:00 236.0 F 0.0
10:00 239.8 F 0.0
15:00 228.4 F 0.0
20:00 231.5 F 0.0
25:00 228.2 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 22:41
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:05
CYCLE COMPLETE

08/25/2022 09:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____