
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006443
10:49 11/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:00
STERILIZING 3:53
136.2 C 221kPa 3:53
136.0 C 221kPa 4:23
136.4 C 225kPa 4:53
136.4 C 222kPa 5:23
136.4 C 223kPa 5:53
136.2 C 222kPa 6:23
136.4 C 224kPa 6:53
136.4 C 222kPa 7:23
136.4 C 222kPa 7:23
VENTING 7:23
AIR DRYING 8:15
CYCLE COMPLETE 8:23

Digital Signature #
0ABA86B2BD902A03
