

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3576

Name: RITTER 1  
Sterilizer ID: V2178198  
07/25/2023 09:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 183.9 F 0.0  
2:00 182.3 F 1.7  
4:00 185.9 F 4.7  
6:00 217.8 F 9.1  
8:00 224.9 F 9.1  
10:00 230.9 F 8.5  
12:00 239.6 F 13.1  
14:00 239.4 F 10.9  
16:00 237.6 F 9.1  
18:00 250.5 F 15.5  
20:00 261.8 F 21.9  
22:00 272.2 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.3 F 28.7  
2:00 271.3 F 28.4  
3:00 272.3 F 28.7  
4:00 271.1 F 28.1

Min 271.1 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
5:00 220.7 F 0.0  
10:00 226.4 F 0.0  
15:00 227.9 F 0.0  
20:00 229.6 F 0.0  
25:00 223.4 F 0.0  
30:00 219.4 F 0.0  
Drying Complete

FILLING: 1:23  
HEATING: 22:29  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:01:22  
CYCLE COMPLETE

07/25/2023 10:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_