
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007508
7:05 07/06/2024

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:16
STERILIZING 4:19
136.4 C 220kPa 4:19
136.1 C 221kPa 4:49
136.4 C 223kPa 5:19
136.4 C 221kPa 5:49
136.4 C 223kPa 6:19
136.0 C 219kPa 6:49
136.0 C 217kPa 7:19
135.8 C 217kPa 7:49
135.8 C 217kPa 7:49
VENTING 7:49
AIR DRYING 8:37
CYCLE COMPLETE 38:37

Digital Signature #
2E56A008905B06C8
