
Midmark M11 - v1.0.0
Total Cycles: 4513

Name: RITTER 1
Sterilizer ID: V2178198
05/01/2024 04:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.7 F 0.0
2:00 158.4 F 2.9
4:00 177.5 F 6.5
6:00 197.5 F 10.6
8:00 211.6 F 4.6
10:00 224.3 F 7.9
12:00 235.7 F 12.4
14:00 235.4 F 9.2
16:00 240.2 F 10.4
18:00 246.5 F 13.4
20:00 257.8 F 19.6
22:00 268.2 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.5 F 29.2
2:00 272.0 F 28.6
3:00 271.3 F 28.5
4:00 272.4 F 29.2

Min 271.2 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 211.8 F 0.0
10:00 216.2 F 0.0
15:00 216.4 F 0.0
20:00 220.7 F 0.0
25:00 224.4 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 23:20
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:02:46
CYCLE COMPLETE

05/01/2024 05:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____