
Midmark M11 - v1.0.0
Total Cycles: 1165

Name: RITTER 1
Sterilizer ID: V2178198
09/13/2021 01:39 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.8 F 0.0
2:00 186.7 F 3.1
4:00 210.3 F 8.3
6:00 224.9 F 13.3
8:00 230.7 F 9.0
10:00 237.6 F 10.2
12:00 243.6 F 12.8
14:00 243.1 F 11.8
16:00 256.1 F 18.6
18:00 267.6 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.7 F 29.0
2:00 271.6 F 28.3
3:00 272.5 F 29.1
4:00 272.4 F 28.7

Min 271.5 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 209.8 F 0.0
10:00 217.4 F 0.0
15:00 207.7 F 0.0
20:00 196.8 F 0.0
25:00 187.2 F 0.0
30:00 180.5 F 0.0
Drying Complete

FILLING: 2:08
HEATING: 19:22
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

09/14/2021 06:27 AM
ACCEPTED: X
REJECTED:
Operator: U01