
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000458
7:27 02/02/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:06
PRESSURIZING 3:08
STERILIZING 4:14
136.0 C 229kPa 4:14
136.2 C 230kPa 4:45
136.4 C 232kPa 5:15
136.0 C 230kPa 5:45
136.4 C 232kPa 6:15
136.4 C 233kPa 6:45
136.4 C 233kPa 7:15
136.4 C 230kPa 7:45
136.4 C 230kPa 7:45
VENTING 7:45
AIR DRYING 8:39
CYCLE COMPLETE 14:21

Digital Signature #
273427BB427DC5F8
