
Midmark M11 - v1.0.0
Total Cycles: 2544

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2022 05:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.9 F 0.0
2:00 197.1 F 4.3
4:00 205.5 F 9.3
6:00 226.4 F 6.5
8:00 231.6 F 9.9
10:00 237.5 F 8.5
12:00 244.7 F 13.3
14:00 244.0 F 12.2
16:00 256.9 F 19.1
18:00 268.5 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.4 F 28.8
2:00 271.4 F 28.4
3:00 272.4 F 28.9
4:00 271.4 F 28.2

Min 271.2 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 235.4 F 0.0
10:00 240.1 F 0.0
15:00 230.7 F 0.0
20:00 226.2 F 0.0
25:00 223.9 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 19:10
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

11/10/2022 06:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____