
Midmark M11 - v1.0.5
Total Cycles: 1255

Name: RITTER2
Sterilizer ID: V2313228
07/05/2022 09:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.6 F 0.0
2:00 191.9 F 3.3
4:00 194.7 F 8.5
6:00 216.1 F 8.8
8:00 228.4 F 9.7
10:00 235.4 F 9.0
12:00 242.8 F 13.4
14:00 243.1 F 12.6
16:00 256.2 F 19.4
18:00 267.7 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.6 F 29.8
2:00 271.6 F 29.3
3:00 272.8 F 29.9
4:00 271.8 F 29.2

Min 271.5 F 29.2
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 209.3 F 0.0
10:00 217.7 F 0.0
15:00 216.8 F 0.0
20:00 215.1 F 0.0
25:00 213.3 F 0.0
30:00 201.3 F 0.0
Drying Complete

FILLING: 2:35
HEATING: 19:19
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

07/05/2022 11:04 AM
ACCEPTED: X
REJECTED:
Operator: U02