
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001206
16:50 03/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:38
STERILIZING 3:48
136.0 C 220kPa 3:48
136.1 C 220kPa 4:18
136.0 C 220kPa 4:48
136.4 C 225kPa 5:18
136.3 C 224kPa 5:48
136.3 C 224kPa 6:18
136.1 C 222kPa 6:48
136.3 C 224kPa 7:18
136.3 C 224kPa 7:18
VENTING 7:18
AIR DRYING 8:10
CYCLE COMPLETE 19:58

Digital Signature #
5AFC9E6BD9E80C7C
