

-----  
Midmark M11 - v1.0.5  
Total Cycles: 833

Name: RITTER2  
Sterilizer ID: V2313228  
02/24/2022 07:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.4 F 0.0  
2:00 76.1 F 0.0  
4:00 86.7 F 0.0  
6:00 124.8 F 0.1  
8:00 161.3 F 3.3  
10:00 181.4 F 6.9  
12:00 219.9 F 10.9  
14:00 222.4 F 5.3  
16:00 229.8 F 9.1  
18:00 240.2 F 12.4  
20:00 239.1 F 10.9  
22:00 235.9 F 8.8  
24:00 249.5 F 15.4  
26:00 261.6 F 22.0  
28:00 272.0 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.4  
1:00 272.0 F 29.7  
2:00 272.2 F 29.9  
3:00 272.5 F 29.9  
4:00 272.3 F 29.9

Min 271.7 F 29.3  
Max 273.2 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.5 F 0.0  
5:00 238.8 F 0.0  
10:00 218.1 F 0.0  
15:00 228.9 F 0.0  
20:00 209.3 F 0.0  
25:00 222.3 F 0.0  
30:00 211.3 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 29:06  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 01:07:10  
CYCLE COMPLETE

02/24/2022 09:02 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02