

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2190

Name: RITTER2  
Sterilizer ID: V2313228  
03/29/2023 07:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 80.5 F 0.0  
2:00 81.2 F 0.0  
4:00 93.6 F 0.0  
6:00 115.4 F 0.4  
8:00 145.1 F 3.5  
10:00 172.2 F 7.2  
12:00 193.8 F 11.1  
14:00 215.1 F 5.0  
16:00 223.7 F 8.8  
18:00 235.1 F 13.3  
20:00 236.1 F 10.2  
22:00 235.9 F 8.7  
24:00 247.1 F 14.5  
26:00 258.5 F 20.8  
28:00 269.3 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.1 F 29.6  
2:00 271.2 F 29.4  
3:00 272.0 F 29.5  
4:00 271.3 F 29.5

Min 271.0 F 28.9  
Max 272.4 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.1 F 0.0  
5:00 205.1 F 0.0  
10:00 213.1 F 0.0  
15:00 223.7 F 0.0  
20:00 227.7 F 0.0  
25:00 225.6 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 29:00  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:07:02  
CYCLE COMPLETE

03/29/2023 08:38 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_