

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1735

Name: RITTER2  
Sterilizer ID: V2313228  
11/16/2022 02:43 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.6 F 0.0  
2:00 187.1 F 2.7  
4:00 207.2 F 8.1  
6:00 232.1 F 10.7  
8:00 234.2 F 10.5  
10:00 232.9 F 8.5  
12:00 238.7 F 10.1  
14:00 249.0 F 15.0  
16:00 261.5 F 22.3  
18:00 272.8 F 29.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.0  
1:00 272.4 F 29.6  
2:00 271.4 F 29.0  
3:00 272.5 F 29.7  
4:00 271.9 F 29.1  
  
Min 271.2 F 28.8  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.2 F 0.0  
5:00 215.1 F 0.0  
10:00 227.8 F 0.0  
15:00 224.2 F 0.0  
20:00 231.6 F 0.0  
25:00 232.2 F 0.0  
30:00 218.3 F 0.0  
Drying Complete

FILLING: 1:48  
HEATING: 18:22  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:24  
CYCLE COMPLETE

11/16/2022 03:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_