

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5398

Name: RITTER 1  
Sterilizer ID: V2178198  
03/07/2025 11:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.7 F 0.0  
2:00 162.4 F 2.2  
4:00 164.8 F 5.6  
6:00 179.9 F 9.6  
8:00 212.6 F 6.7  
10:00 218.7 F 6.8  
12:00 232.7 F 11.4  
14:00 231.5 F 7.6  
16:00 243.5 F 13.1  
18:00 241.3 F 11.4  
20:00 253.4 F 17.4  
22:00 264.0 F 23.8  
24:00 271.5 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.7  
1:00 272.3 F 29.5  
2:00 271.7 F 28.9  
3:00 271.7 F 29.2  
4:00 272.8 F 30.0

Min 271.0 F 28.6  
Max 273.3 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 204.7 F 0.0  
10:00 204.5 F 0.0  
15:00 213.5 F 0.0  
20:00 225.5 F 0.0  
25:00 230.3 F 0.0  
30:00 227.7 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 24:08  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:02:52  
CYCLE COMPLETE

03/07/2025 12:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_