
Midmark M11 - v1.0.0
Total Cycles: 3894

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2023 10:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.8 F 0.0
2:00 172.2 F 3.3
4:00 190.2 F 7.6
6:00 227.6 F 12.8
8:00 228.7 F 7.4
10:00 238.3 F 12.7
12:00 237.8 F 9.7
14:00 236.6 F 8.6
16:00 248.5 F 14.4
18:00 260.4 F 21.0
20:00 271.1 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.5 F 28.7
2:00 272.9 F 29.1
3:00 271.5 F 28.7
4:00 272.6 F 28.8

Min 271.4 F 28.4
Max 272.9 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.8 F 0.0
5:00 249.0 F 0.0
10:00 242.3 F 0.0
15:00 226.0 F 0.0
20:00 220.9 F 0.0
25:00 223.7 F 0.0
30:00 231.3 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 21:23
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:00:17
CYCLE COMPLETE

10/12/2023 11:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____